

Pos.	Descrizione	Description	Codice Code	Q.tà Q.ty
1	Pompa per vuoto lubrificata LC 40WR	Oil lubricated vacuum pump LC 40WR	9690036 (*)	1
2	Riduzione 1-1/2"G M / 1"G F	Fitting 1-1/2"G M / 1"G F	2001009	1
3	Curva 1"G M/F	Elbow 1"G M/F	2101017	2
4	Tubo 1"Gx60	Pipe 1"Gx60	5001011	1
5	Valvola VNR.1	Non-return valve VNR.1	9007103	1
6	Tubo 1"Gx80	Pipe 1"Gx80	5001015	1
7	Dado 1"G	Nut 1"G	0202007	2
8	Filtro separatore di condensa FSC.60	Condensation separator filter FSC.60	9001029	1
9	Nipples 1/4"G	Fitting 1/4"G	2201001	1
10	Raccordo a "T" F/M/F 1/4"G	"T" fitting F/M/F 1/4"G	2106048	1
11	Dado 1/4"G	Nut 1/4"G	0202002	1
12	Vuotometro VMR.63	Vacuum gauge VMR.63	9009009	1
13	Rubinetto a sfera F/F 1/4"G	Valve F/F 1/4"G	2107001	1
14	Fascetta acciaio 24-36	Steel clamp 24-36	1201008	2
15	Raccordo portagomma Ø25x1"G	Hose fitting Ø25x1"G	2103012	2
16	Tubo armovin Ø25xØ33 L=630	Armovin pipe Ø25xØ33 L=630	5004006	1
17	Rubinetto a sfera F/F 1"G	Valve F/F 1"G	2107008	1
18	Nipples M/M 1"G foro 1/4"G ottone giallo	Yellow brass fitting M/M 1"G hole 1/4"G	2201020	1
19	Tappo 1/4"G + OR	Knob 1/4"G + O-ring	1110002	1
20	Vite TE M8x25	Screw hexagonal head M8x25	0102024	8
21	Rondella piana M8	Washer M8	0301005	11
22	Raccordo portagomma curvo Ø25x1"G	Elbow hose fitting Ø25x1"G	2103119	1
23	Serbatoio 25Lt.	25litre tank	5601043	1
24	Guarnizione adesiva 10x3x1600	Adhesive gasket 10x3 L=1600	4903001	1
25	Coperchio serbatoio 25Lt.	25litre tank cover	5601044	1
26	Indicatore elettrico-visivo del livello dei liquidi	Electrical visual liquid level gauge	9013004	1
27	Nipples 1/2"G	Fitting 1/2"G	2201002	1
28	Valvola VNR.1/2"G	Non-return valve VNR 1/2"G	9007101	1
29	Rondella groower Ø8	Groower washer Ø8	0303002	3
30	Vite TE M8x12	Screw hexagonal head M8x12	0102042	3

(*) In base alla tensione di funzionamento della pompa, l'estensione del codice tecnico potrebbe subire delle variazioni

The suffix of the product reference code changes according to the voltage at which the pump operates

Subject to change without prior notice

Subject to change without prior notice

